


RSO Aurora

Sample ID: 1903CSALA2244.7790
Strain: RSO Aurora
Matrix: Concentrates & Extracts
Type: Other
Sample Size: 4 units; Batch:

Produced: N/A
Collected: 03/28/2019
Received: 03/28/2019
Completed: 04/03/2019
Batch#:

Client:
Ricksimpsonoil California
Lic. #
Murrieta, CA 92562



Cannabinoids

Testing method: HPLC-SOP 101

Complete

| Analyte | LOD % | LOQ % | Mass % | Mass mg/g |
|--------------|----------|----------|--------------|--------------|
| THCa | 0.00 | 0.00 | ND | ND |
| Δ9-THC | 0.00 | 0.00 | 76.50 | 765.0 |
| Δ8-THC | 0.00 | 0.00 | ND | ND |
| THCV | 0.00 | 0.00 | 0.57 | 5.7 |
| CBDa | 0.00 | 0.00 | ND | ND |
| CBD | 0.00 | 0.00 | ND | ND |
| CBDV | 0.00 | 0.00 | ND | ND |
| CBN | 0.00 | 0.00 | 0.58 | 5.8 |
| CBGa | 0.00 | 0.00 | ND | ND |
| CBG | 0.00 | 0.00 | 1.56 | 15.6 |
| CBC | 0.00 | 0.00 | 1.06 | 10.6 |
| Total | | | 80.27 | 802.7 |

 $Total\ THC = THCa * 0.877 + d9-THC$
 $Total\ CBD = CBDa * 0.877 + CBD$

LOQ = Limit of Quantitation; LOD = Limit of Detection; NR = Not Reported; NT = Not Tested; ND = Not Detected. The reported result is based on a sample weight with the applicable moisture content for that sample;


 ISO / IEC 17025:2017 ACCREDITED LABORATORY
 Accreditation No. 73653

 Mallory Speakman
 Laboratory Director - LA
 04/03/2019

 Anya Engen
 Quality Manager
 04/03/2019




RSO Aurora

Sample ID: 1903CSALA2244.7790
Strain: RSO Aurora
Matrix: Concentrates & Extracts
Type: Other
Sample Size: 4 units; Batch:

Produced: N/A
Collected: 03/28/2019
Received: 03/28/2019
Completed: 04/03/2019
Batch#:

Client
Ricksimpsonoil California
Lic. #
Murrieta, CA 92562

Residual Solvents

Testing method: HSGCMS-SOP 202

| Analyte | LOD | LOQ | Limit | Mass | Status |
|---------------------|--------|--------|-------|------|--------|
| | µg/g | µg/g | µg/g | µg/g | |
| 1,2-Dichloro-Ethane | 0.024 | 0.24 | 1 | ND | Pass |
| Acetone | 0.112 | 1.12 | 5000 | ND | Pass |
| Acetonitrile | 0.0196 | 0.196 | 410 | ND | Pass |
| Benzene | 0.0096 | 0.096 | 1 | ND | Pass |
| Butane | 0.2 | 2 | 5000 | ND | Pass |
| Chloroform | 0.0288 | 0.288 | 1 | ND | Pass |
| Ethanol | 0.112 | 1.12 | 5000 | ND | Pass |
| Ethyl-Acetate | 0.112 | 1.12 | 5000 | ND | Pass |
| Ethyl-Ether | 0.112 | 1.12 | 5000 | ND | Pass |
| Ethylene Oxide | 0.024 | 0.24 | 1 | ND | Pass |
| Heptane | 0.112 | 1.12 | 5000 | ND | Pass |
| Isopropanol | 0.112 | 1.12 | 5000 | ND | Pass |
| Methanol | 0.144 | 1.44 | 3000 | ND | Pass |
| Methylene-Chloride | 0.0288 | 0.288 | 1 | ND | Pass |
| n-Hexane | 0.1392 | 1.392 | 290 | ND | Pass |
| Pentane | 0.112 | 1.12 | 5000 | ND | Pass |
| Propane | 0.08 | 0.8 | 5000 | ND | Pass |
| Toluene | 0.4272 | 4.272 | 890 | ND | Pass |
| Trichloroethene | 0.0384 | 0.384 | 1 | ND | Pass |
| Xylenes | 1.0416 | 10.416 | 2170 | ND | Pass |

Date Tested: 04/02/2019
LOQ = Limit of Quantitation; LOD = Limit of Detection; NT = Not Tested; ND = Not Detected.



ISO / IEC 17025:2017 ACCREDITED LABORATORY
Accreditation No. 73653

Mallory Speakman
Laboratory Director - LA
04/03/2019

Anya Engen
Quality Manager
04/03/2019

